



Högloms Lego Montage AB
Cedersdalsvägen 7
Box 59
S-18631 Vallentuna
NAT 08-51430670 INT +46-8-51430670

Calibration Protocol

PFM Serial No: 1325
Static Pressure 0 kPa

Reference		Reading	
0,00	mbar	0	mbar
100,00	mbar	99,7	mbar
200,00	mbar	199,9	mbar
300,00	mbar	299,8	mbar
400,00	mbar	400	mbar
500,00	mbar	499,5	mbar
600,00	mbar	599,7	mbar
700,00	mbar	700	mbar
800,00	mbar	800,5	mbar
900,00	mbar	900,5	mbar
1000,00	mbar	1000,3	mbar

This calibration is done with Pressure balance
Type T2300-1
Serial No 4706-87
Range 0,1-35 bar

This calibration is done and approved by:

Janne Hämäläinen

Date:

081205



1325

